

SPECIFICATION FOR DIATHERMY MACHINE WITH BIPOLAR ARRANGEMENTS MONOPOLAR ARRANGEMENTS WITH FOOT AND HAND OPERATION.

- Monopolar - Cutting, coagulation and blend contact coagulation.
- Bipolar coagulation and special tripolar cutting.
- Should have facility for under water cutting.
- Should have both hand and foot switch operation.
- Should have audio sound for cutting and coagulation mode.
- Should have Neutral Electrode safety system to prevent burns.
- Should have all Cables and Accessories.