

### **SPECIFICATION FOR INFUSION PUMP**

- Should have a flow rate of 1 – 1000ml/hour in normal mode and 0.1 – 100 ml/hr with 0.1 ml increment in micro infusion mode.
- Should have special modes like Flow rate and time, Primary and secondary, pump up & pump down etc.
- Should have facility to display and enter drug name.
- Key board locking facility to protect against change of settings.
- Should have alarm facility for low battery, occlusion, air bubble, end of infusion, line disconnection etc.
- Should be complete with all cables and accessories.