

SPECIFICATION FOR ANESTHESIA MACHINE WITH VENTILATOR

- **Integrated Ventilator.**
- **Volume Controlled Ventilation.**
- **Integrated Compact breathing system.**
- **Latex free and autoclavable bellows.**
- **Compact autoclavable soda lime canister.**
- **Color LCD Display.**
- **Dual cascaded flow meters (O₂, N₂O) with mechanical hypoxic guard.**
- **Audio visual alarms.**
- **Pressure, volume and FiO₂% monitoring.**
- **Airway Pressure gauge, Airway pressure and Flow Waveforms.**
- **Monitor shelf, At least two drawers.**
- **Selectatec mounting of up to 2 vaporizers, at least 1 vaporizer (Tec 5) should be provided either sevoflurane or isoflurane.**
- **Pull Handle.**
- **1 Adult Bains circuit and 1 Pediatric. JACKSON REE's circuit should be provided**